



Alternate beamsplitter design

Locations of eyepiece and light source are interchangeable.  
 The reticle at the light source is typically a Siemen's Star.  
 The reticle at the eyepiece may be a simple circle, or a crosshair.  
 The objective should be a well corrected achromat.  
 The focal length should be equal or longer than the lens under test.  
 A pair of glued prisms is often used for the beamsplitter.